



MANUFACTURER'S EU DECLARATION OF CONFORMITY

This declaration is made under the sole responsibility of the manufacturer under Annexe IV of the Medical Device Regulation (EU) 2017/745

Manufacturer: Lancaster Healthcare Ltd t/a Flexyfoot, Unit 8 Brook Mill Industrial Estate, Wrea Green, Preston, Lancashire, PR4 2PH +44 1483722686 sales@flexyfoot.com

Authorised Representative: CareServe France SAS, 847 route de Frans, Bâtiment J, 69400 VILLEFRANCHE SUR SAONE. +33 482835000 gilles.adam@careserve.fr

EUDAMED SRN: Not currently available

Medical Devices(s): Flexyfoot Walking Sticks.

OVAL HANDLE/COLOUR

	SKU	UDI-DI
Telescopic Walking Stick/Cane - Black	FS-301-BLACK	5060219050802
Telescopic Walking Stick/Cane - Blue	FS-301-BLUE	5060219050819
Telescopic Walking Stick/Cane - Orange	FS-301-ORANGE	5060219051311
Telescopic Walking Stick/Cane - Red	FS-301-RED	5060219051328
Folding Walking Stick/Cane - Black	FS-302B-BLACK	5060219050901
Folding Walking Stick/Cane - Blue	FS-302B-BLUE	5060219050918
Folding Walking Stick/Cane - Orange	FS-302B-ORANGE	5060219051335
Folding Walking Stick/Cane - Red	FS-302B-RED	5060219051342

CORK HANDLE/COLOUR

	SKU	UDI-DI
Telescopic Walking Stick/Cane - Black	FSC-301-BLACK	5060219050864
Telescopic Walking Stick/Cane - Blue	FSC-301-BLUE	5060219050871
Telescopic Walking Stick/Cane - Orange	FSC-301-ORANGE	5060219051359
Telescopic Walking Stick/Cane - Red	FSC-301-RED	5060219051366
Folding Walking Stick/Cane - Black	FSC-302B-BLACK	5060219050925
Folding Walking Stick/Cane - Blue	FSC-302B-BLUE	5060219050932
Folding Walking Stick/Cane - Orange	FSC-302B-ORANGE	5060219051373
Folding Walking Stick/Cane - Red	FSC-302B-RED	5060219051380

DERBY HANDLE/COLOUR

	SKU	UDI-DI
Telescopic Walking Stick/Cane - Black	FSD-301-BLACK	5060219050017
Telescopic Walking Stick/Cane - Blue	FSD-301-BLUE	5060219050024
Telescopic Walking Stick/Cane - Orange	FSD-301-ORANGE	5060219051397
Telescopic Walking Stick/Cane - Red	FSD-301-RED	5060219051403
Folding Walking Stick/Cane - Black	FSD-302B-BLACK	5060219050031
Folding Walking Stick/Cane - Blue	FSD-302B-BLUE	5060219050048
Folding Walking Stick/Cane - Orange	FSD-302B-ORANGE	5060219051410
Folding Walking Stick/Cane - Red	FSD-302B-RED	5060219051427

Classification: Class 1, Non-invasive Device under Annexe VIII, Chapter III, 4. Non-invasive devices, Rule 1

GMDN Code & Term: 31112 Single-base walking stick

Scope of application: A rod, typically made of wood or metal, with one leg/foot and a handgrip or curved handle and without forearm support used to assist a person with a disability to maintain their balance and support their weight while walking. It may be of fixed or adjustable length, and collapsible or non-collapsible.

Notified Body: N/A

Standards Applied: **BS EN ISO 24415-1:2009** Tips for assistive products for walking – Requirements and Test Methods. Part 1: Friction of tips

BS EN ISO 24415-2:2011 Tips for assistive products for walking – Requirements and Test Methods. Part 2: Durability of tips for crutches

BS EN 1985:1999: Walking aids – General requirements and test methods - Clause 7 only - Static load test

Date of Issue: 01st July 2021

Authorised Signatory:



Georgie Powell
Managing Director
On behalf of Lancaster Healthcare Ltd